

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100125\
 Data File : VX047947.D
 Acq On : 01 Oct 2025 17:35
 Operator : JC/MD
 Sample : Q3194-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT188S1-20250922

Quant Time: Oct 03 01:27:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	123198	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	253162	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	260063	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	124890	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	103869	52.967	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.940%	
35) Dibromofluoromethane	5.379	113	85821	50.547	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.100%	
50) Toluene-d8	8.635	98	291465	49.612	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.220%	
62) 4-Bromofluorobenzene	11.067	95	120298	53.954	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.900%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1438	1.687	ug/l #	59
44) Trichloroethene	7.110	130	10431	5.726	ug/l	95
64) Tetrachloroethene	9.263	164	571	0.337	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100125\
Data File : VX047947.D
Acq On : 01 Oct 2025 17:35
Operator : JC/MD
Sample : Q3194-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT188S1-20250922

Quant Time: Oct 03 01:27:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

