

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100125\
 Data File : VX047948.D
 Acq On : 01 Oct 2025 17:56
 Operator : JC/MD
 Sample : Q3193-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW10-MW01D-20250923

Quant Time: Oct 03 01:27:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	114654	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	240635	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	251538	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	129301	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	109782	60.154	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 120.300%#		
35) Dibromofluoromethane	5.379	113	89966	55.747	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 111.500%		
50) Toluene-d8	8.635	98	302786	54.222	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 108.440%		
62) 4-Bromofluorobenzene	11.067	95	130975	61.801	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 123.600%#		
Target Compounds						
						Qvalue
16) Acetone	2.374	43	2236	2.818	ug/l #	73
44) Trichloroethene	7.123	130	434	0.251	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100125\
Data File : VX047948.D
Acq On : 01 Oct 2025 17:56
Operator : JC/MD
Sample : Q3193-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW10-MW01D-20250923

Quant Time: Oct 03 01:27:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

