

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100218\
 Data File : VX004916.D
 Acq On : 02 Oct 2018 13:11
 Operator : JC/MD
 Sample : VX1002WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1002WBL01

Quant Time: Oct 03 08:14:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X100218W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Oct 03 04:49:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.02	128	63323	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.86	114	334311	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	301105	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.07	65	139717	29.04	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	96.80%
60) 4-Bromofluorobenzene	11.14	95	132609	26.50	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	88.33%
63) Toluene-d8	8.71	98	401864	29.90	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	99.67%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100218\
Data File : VX004916.D
Acq On : 02 Oct 2018 13:11
Operator : JC/MD
Sample : VX1002WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1002WBL01

Quant Time: Oct 03 08:14:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X100218W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Oct 03 04:49:12 2018
Response via : Initial Calibration

