

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100218\
 Data File : VX004915.D
 Acq On : 02 Oct 2018 12:33
 Operator : JC/MD
 Sample : VSTDICV020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX100218

Quant Time: Oct 03 07:56:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X100218W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Oct 03 04:49:12 2018
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	85	0.00
2 M	Dichlorodifluoromethane	2.336	2.777	-18.9	99	0.00
3 M	Chloromethane	2.423	2.900	-19.7	100	0.00
4 M	Vinyl Chloride	2.477	2.885	-16.5	99	0.00
5 M	Bromomethane	1.189	1.314	-10.5	103	0.00
6 M	Chloroethane	1.422	1.622	-14.1	96	0.00
7 M	Trichlorofluoromethane	3.156	3.693	-17.0	100	0.00
8 T	Diethyl Ether	1.286	1.465	-13.9	99	0.00
9	1,1,2-Trichlorotrifluoroeth	1.812	2.145	-18.4	102	0.00
10 M	1,1-Dichloroethene	1.717	1.938	-12.9	98	0.00
11	Methyl Iodide	1.398	1.216	13.0	103	0.00
12	Methyl Acetate	3.462	4.038	-16.6	98	0.00
13 M	Acrolein	0.320	0.343	-7.2	96	0.00
14 M	Acrylonitrile	1.415	1.602	-13.2	97	0.00
15 M	Acetone	0.473	0.515	-8.9	85	0.00
16 M	Carbon Disulfide	5.035	5.929	-17.8	102	0.00
17	Allyl chloride	3.817	4.353	-14.0	97	0.00
18 M	Methylene Chloride	1.996	2.259	-13.2	98	0.00
19 M	trans-1,2-Dichloroethene	1.834	2.105	-14.8	100	0.00
20 T	Diisopropyl ether	6.509	7.774	-19.4	106	0.00
21 M	1,1-Dichloroethane	3.586	4.152	-15.8	98	0.00
22 M	cis-1,2-Dichloroethene	2.029	2.286	-12.7	99	0.00
23 M	tert-Butyl Alcohol	0.653	0.781	-19.6	98	0.00
24 M	Methyl tert-Butyl Ether	6.308	7.143	-13.2	98	0.00
25 M	Chloroform	3.375	3.788	-12.2	98	0.00
26	Cyclohexane	3.078	3.539	-15.0	100	0.00
27 s	1,2-Dichloroethane-d4	2.280	2.283	-0.1	88	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
29	1,1-Dichloropropene	0.444	0.511	-15.1	100	0.00
30 M	2-Butanone	0.361	0.404	-11.9	92	0.00
31	2,2-Dichloropropane	0.487	0.564	-15.8	100	0.00
32 M	1,1,1-Trichloroethane	0.510	0.571	-12.0	96	0.00
33 M	Carbon Tetrachloride	0.448	0.499	-11.4	97	0.00
34 M	Benzene	1.347	1.526	-13.3	97	0.00
35	Methacrylonitrile	0.335	0.378	-12.8	96	0.00
36 M	1,2-Dichloroethane	0.485	0.549	-13.2	96	0.00
37 M	Trichloroethene	0.355	0.410	-15.5	101	0.00
38	Methylcyclohexane	0.527	0.614	-16.5	106	0.00
39 M	1,2-Dichloropropane	0.359	0.408	-13.6	99	0.00
40	Dibromomethane	0.233	0.264	-13.3	99	0.00
41 M	Bromodichloromethane	0.448	0.495	-10.5	97	0.00
42 M	Vinyl Acetate	1.044	1.281	-22.7	106	0.00
43	Ethyl Acetate	0.595	0.687	-15.5	100	0.00
44	Isopropyl Acetate	0.888	0.979	-10.2	96	0.00
45 T	1,4-Dioxane	0.010	0.012#	-20.0	94	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100218\
 Data File : VX004915.D
 Acq On : 02 Oct 2018 12:33
 Operator : JC/MD
 Sample : VSTDICV020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX100218

Quant Time: Oct 03 07:56:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X100218W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Oct 03 04:49:12 2018
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.461	0.506	-9.8	96	0.00
47	n-amyl Acetate	0.820	0.855	-4.3	96	0.00
48 M	t-1,3-Dichloropropene	0.513	0.561	-9.4	98	0.00
49 T	cis-1,3-Dichloropropene	0.556	0.623	-12.1	97	0.00
50 M	1,1,2-Trichloroethane	0.352	0.399	-13.4	98	0.00
51	Ethyl methacrylate	0.559	0.611	-9.3	99	0.00
52	1,3-Dichloropropane	0.594	0.663	-11.6	95	0.00
53 M	Dibromochloromethane	0.365	0.394	-7.9	97	0.00
54 M	1,2-Dibromoethane	0.367	0.409	-11.4	96	0.00
55 M	2-Chloroethyl vinyl ether	0.295	0.323	-9.5	98	0.00
56 M	Bromoform	0.305	0.317	-3.9	97	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
58 M	4-Methyl-2-Pentanone	0.674	0.751	-11.4	95	0.00
59 M	2-Hexanone	0.555	0.606	-9.2	94	0.00
60 S	4-Bromofluorobenzene	0.499	0.500	-0.2	91	0.00
61 M	Tetrachloroethene	0.365	0.426	-16.7	99	0.00
62 M	Toluene	1.577	1.778	-12.7	98	0.00
63 S	Toluene-d8	1.339	1.309	2.2	84	0.00
64 M	Chlorobenzene	1.018	1.143	-12.3	98	0.00
65	1,1,1,2-Tetrachloroethane	0.370	0.398	-7.6	94	0.00
66 M	Ethyl Benzene	1.721	1.917	-11.4	99	0.00
67 M	m/p-Xylenes	0.670	0.745	-11.2	98	0.00
68 M	o-Xylene	0.646	0.717	-11.0	99	0.00
69 M	Styrene	1.086	1.189	-9.5	98	0.00
70	Isopropylbenzene	1.730	1.943	-12.3	98	0.00
71 M	1,1,2,2-Tetrachloroethane	0.643	0.689	-7.2	97	0.00
72	1,2,3-Trichloropropane	0.585	0.638	-9.1	95	0.00
73	Bromobenzene	0.476	0.516	-8.4	97	0.00
74	n-propylbenzene	2.035	2.296	-12.8	100	0.00
75	2-Chlorotoluene	1.241	1.382	-11.4	99	0.00
76	1,3,5-Trimethylbenzene	1.493	1.670	-11.9	99	0.00
77	t-1,4-Dichloro-2-butene	0.192	0.189	1.6	98	0.00
78	4-Chlorotoluene	1.443	1.588	-10.0	99	0.00
79	tert-butylbenzene	1.584	1.744	-10.1	99	0.00
80	1,2,4-Trimethylbenzene	1.522	1.693	-11.2	98	0.00
81	sec-Butylbenzene	1.762	1.982	-12.5	101	0.00
82	p-Isopropyltoluene	1.584	1.744	-10.1	99	0.00
83 M	1,3-Dichlorobenzene	0.879	0.954	-8.5	98	0.00
84 M	1,4-Dichlorobenzene	0.892	0.973	-9.1	99	0.00
85	n-Butylbenzene	1.439	1.583	-10.0	102	0.00
86 T	Hexachloroethane	0.261	0.272	-4.2	99	0.00
87 M	1,2-Dichlorobenzene	0.894	0.948	-6.0	97	0.00
88	1,2-Dibromo-3-Chloropropane	0.148	0.153	-3.4	93	0.00
89	1,2,4-Trichlorobenzene	0.667	0.728	-9.1	103	0.00
90	Hexachlorobutadiene	0.329	0.382	-16.1	109	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100218\
Data File : VX004915.D
Acq On : 02 Oct 2018 12:33
Operator : JC/MD
Sample : VSTDICV020
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX100218

Quant Time: Oct 03 07:56:34 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X100218W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Oct 03 04:49:12 2018
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	2.073	2.200	-6.1	97	0.00
92	1,2,3-Trichlorobenzene	0.684	0.731	-6.9	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0