

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100219\
 Data File : VX012757.D
 Acq On : 02 Oct 2019 13:42
 Operator : JC/SP
 Sample : K5082-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-DUP05-20190926

Quant Time: Oct 03 03:11:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	178594	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	315836	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	266478	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	100515	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	126519	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
35) Dibromofluoromethane	5.49	113	95765	50.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.82%	
50) Toluene-d8	8.71	98	367575	50.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.48%	
62) 4-Bromofluorobenzene	11.14	95	121736	42.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.48%	
Target Compounds						
16) Acetone	2.43	43	8004	5.810	ug/l	96
30) Chloroform	5.20	83	1081	0.264	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100219\
Data File : VX012757.D
Acq On : 02 Oct 2019 13:42
Operator : JC/SP
Sample : K5082-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-DUP05-20190926

Quant Time: Oct 03 03:11:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

