

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100219\
 Data File : VX012765.D
 Acq On : 02 Oct 2019 17:13
 Operator : JC/SP
 Sample : K5159-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-191001

Quant Time: Oct 03 04:06:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	170328	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	297158	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	244949	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	90093	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	120036	48.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.80%	
35) Dibromofluoromethane	5.49	113	90968	51.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.80%	
50) Toluene-d8	8.71	98	348512	51.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.28%	
62) 4-Bromofluorobenzene	11.14	95	109915	41.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.04%	
Target Compounds						
3) Chloromethane	1.32	50	821	0.373	ug/l #	77
16) Acetone	2.43	43	1828	1.391	ug/l	90
27) cis-1,2-Dichloroethene	4.59	96	8439	3.585	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100219\
Data File : VX012765.D
Acq On : 02 Oct 2019 17:13
Operator : JC/SP
Sample : K5159-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-191001

Quant Time: Oct 03 04:06:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

