

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100219\
 Data File : VX012768.D
 Acq On : 02 Oct 2019 18:46
 Operator : JC/SP
 Sample : K5157-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0539

Quant Time: Oct 03 05:01:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	172576	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	299167	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	255138	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	97864	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	119571	47.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.16%	
35) Dibromofluoromethane	5.50	113	89972	50.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.00%	
50) Toluene-d8	8.71	98	348290	50.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.52%	
62) 4-Bromofluorobenzene	11.14	95	117147	43.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.84%	
Target Compounds						
4) Vinyl Chloride	1.40	62	5338	2.409	ug/l #	78
16) Acetone	2.43	43	3423	2.572	ug/l	92
21) trans-1,2-Dichloroethene	3.16	96	4151	2.019	ug/l	86
27) cis-1,2-Dichloroethene	4.59	96	4563	1.913	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100219\
Data File : VX012768.D
Acq On : 02 Oct 2019 18:46
Operator : JC/SP
Sample : K5157-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-0539

Quant Time: Oct 03 05:01:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

