

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
 Data File : VX043217.D
 Acq On : 02 Oct 2024 09:28
 Operator : JC/MD
 Sample : VX1002MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1002MBL01

Quant Time: Oct 02 22:56:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	28252	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	56680	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	47631	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	15623	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	27221	58.997	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	118.000%
35) Dibromofluoromethane	5.385	113	21421	54.801	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.600%
50) Toluene-d8	8.647	98	64660	47.733	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.460%
62) 4-Bromofluorobenzene	11.079	95	20862	42.392	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	84.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
Data File : VX043217.D
Acq On : 02 Oct 2024 09:28
Operator : JC/MD
Sample : VX1002MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1002MBL01

Quant Time: Oct 02 22:56:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

