

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
 Data File : VX043218.D
 Acq On : 02 Oct 2024 09:51
 Operator : JC/MD
 Sample : VX1002WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1002WBL01

Quant Time: Oct 02 22:57:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	86839	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	164111	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	144102	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	61638	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	73459	51.797	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.600%
35) Dibromofluoromethane	5.379	113	55038	48.630	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.260%
50) Toluene-d8	8.647	98	197924	50.464	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.920%
62) 4-Bromofluorobenzene	11.079	95	72287	50.732	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.460%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
Data File : VX043218.D
Acq On : 02 Oct 2024 09:51
Operator : JC/MD
Sample : VX1002WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1002WBL01

Quant Time: Oct 02 22:57:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

