

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
 Data File : VX043221.D
 Acq On : 02 Oct 2024 11:11
 Operator : JC/MD
 Sample : PB163826TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB163826TB

Quant Time: Oct 02 22:59:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	87545	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	169367	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	152078	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67689	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	75565	52.852	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.700%			
35) Dibromofluoromethane	5.379	113	57593	49.308	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.620%			
50) Toluene-d8	8.647	98	203932	50.382	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.760%			
62) 4-Bromofluorobenzene	11.079	95	75901	51.615	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 103.240%			
Target Compounds						
16) Acetone	2.386	43	26077	43.675	ug/l	100
18) Methyl Acetate	2.709	43	1630	1.088	ug/l #	83
20) Methylene Chloride	2.788	84	5795	5.264	ug/l	93
43) Isopropyl Acetate	6.342	43	89742	30.662	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
Data File : VX043221.D
Acq On : 02 Oct 2024 11:11
Operator : JC/MD
Sample : PB163826TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB163826TB

Quant Time: Oct 02 22:59:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

