

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
 Data File : VX043225.D
 Acq On : 02 Oct 2024 12:43
 Operator : JC/MD
 Sample : PB163826ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB163826ZHE#04

Quant Time: Oct 02 23:01:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	84081	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	161375	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	135591	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57429	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72207	52.584	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	105.160%	
35) Dibromofluoromethane	5.385	113	53861	48.397	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	96.800%	
50) Toluene-d8	8.647	98	189983	49.260	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	98.520%	
62) 4-Bromofluorobenzene	11.079	95	66635	47.558	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	95.120%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	26027	45.387	ug/l	98
18) Methyl Acetate	2.715	43	2338	1.625	ug/l #	89
20) Methylene Chloride	2.788	84	4035	3.816	ug/l	99
43) Isopropyl Acetate	6.342	43	87806	31.486	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
Data File : VX043225.D
Acq On : 02 Oct 2024 12:43
Operator : JC/MD
Sample : PB163826ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB163826ZHE#04

Quant Time: Oct 02 23:01:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

