

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
 Data File : VX043227.D
 Acq On : 02 Oct 2024 13:29
 Operator : JC/MD
 Sample : PB163826ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB163826ZHE#06

Quant Time: Oct 03 05:03:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	86810	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	164518	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	136598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56822	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	75280	53.099	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.200%			
35) Dibromofluoromethane	5.385	113	55843	49.219	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.440%			
50) Toluene-d8	8.647	98	192066	48.849	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 97.700%			
62) 4-Bromofluorobenzene	11.079	95	65884	46.124	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 92.240%			
Target Compounds						
16) Acetone	2.386	43	14830	25.048	ug/l	91
18) Methyl Acetate	2.709	43	2675	1.801	ug/l	97
20) Methylene Chloride	2.788	84	11985	10.979	ug/l	93
25) 2-Butanone	4.586	43	4314	5.314	ug/l	93
43) Isopropyl Acetate	6.342	43	97761	34.386	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
Data File : VX043227.D
Acq On : 02 Oct 2024 13:29
Operator : JC/MD
Sample : PB163826ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB163826ZHE#06

Quant Time: Oct 03 05:03:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

