

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
 Data File : VX043228.D
 Acq On : 02 Oct 2024 13:59
 Operator : JC/MD
 Sample : PB163826ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB163826ZHE#07

Quant Time: Oct 02 23:02:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	90363	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	172530	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153068	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69498	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	76999	52.176	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.360%			
35) Dibromofluoromethane	5.379	113	57369	48.216	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.440%			
50) Toluene-d8	8.647	98	207360	50.289	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.580%			
62) 4-Bromofluorobenzene	11.079	95	77659	51.843	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 103.680%			
Target Compounds						
16) Acetone	2.379	43	16569	26.885	ug/l	98
18) Methyl Acetate	2.703	43	1804	1.167	ug/l	95
20) Methylene Chloride	2.788	84	12096	10.645	ug/l	95
43) Isopropyl Acetate	6.336	43	103218	34.619	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
Data File : VX043228.D
Acq On : 02 Oct 2024 13:59
Operator : JC/MD
Sample : PB163826ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB163826ZHE#07

Quant Time: Oct 02 23:02:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

