

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
 Data File : VX043229.D
 Acq On : 02 Oct 2024 14:22
 Operator : JC/MD
 Sample : PB163826ZHE#08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB163826ZHE#08

Quant Time: Oct 03 05:07:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	82582	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	164388	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	145047	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62002	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72444	53.714	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.420%			
35) Dibromofluoromethane	5.385	113	53751	47.413	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 94.820%			
50) Toluene-d8	8.647	98	201451	51.276	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.560%			
62) 4-Bromofluorobenzene	11.079	95	69918	48.987	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 97.980%			
Target Compounds						
16) Acetone	2.386	43	14747	26.183	ug/l	97
18) Methyl Acetate	2.709	43	2774	1.963	ug/l	99
20) Methylene Chloride	2.788	84	11428	11.005	ug/l	99
25) 2-Butanone	4.586	43	4363	5.650	ug/l #	89
43) Isopropyl Acetate	6.342	43	96422	33.942	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
Data File : VX043229.D
Acq On : 02 Oct 2024 14:22
Operator : JC/MD
Sample : PB163826ZHE#08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB163826ZHE#08

Quant Time: Oct 03 05:07:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

