

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
 Data File : VX043232.D
 Acq On : 02 Oct 2024 15:31
 Operator : JC/MD
 Sample : P4221-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-WATER-REF-4

Quant Time: Oct 02 23:04:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	86471	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	169288	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	152576	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66517	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	75198	53.249	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.500%
35) Dibromofluoromethane	5.385	113	55933	47.909	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.820%
50) Toluene-d8	8.647	98	204476	50.540	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.080%
62) 4-Bromofluorobenzene	11.079	95	75659	51.475	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.940%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
Data File : VX043232.D
Acq On : 02 Oct 2024 15:31
Operator : JC/MD
Sample : P4221-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-WATER-REF-4

Quant Time: Oct 02 23:04:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

