

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
 Data File : VX043233.D
 Acq On : 02 Oct 2024 15:55
 Operator : JC/MD
 Sample : P4221-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Oct 03 02:34:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	86279	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	168784	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153866	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63571	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	75542	53.611	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.220%
35) Dibromofluoromethane	5.385	113	55941	48.059	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.120%
50) Toluene-d8	8.647	98	203819	50.528	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.060%
62) 4-Bromofluorobenzene	11.079	95	73010	49.821	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
Data File : VX043233.D
Acq On : 02 Oct 2024 15:55
Operator : JC/MD
Sample : P4221-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Oct 03 02:34:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

