

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
 Data File : VX043235.D
 Acq On : 02 Oct 2024 16:41
 Operator : JC/MD
 Sample : P4232-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-R

Quant Time: Oct 03 04:48:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	84801	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	164859	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	146230	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66066	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	75200	54.299	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.600%			
35) Dibromofluoromethane	5.379	113	55096	48.460	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.920%			
50) Toluene-d8	8.647	98	200229	50.820	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.640%			
62) 4-Bromofluorobenzene	11.079	95	72154	50.409	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 100.820%			
Target Compounds						
16) Acetone	2.386	43	40227	69.554	ug/l	98
18) Methyl Acetate	2.709	43	1775	1.223	ug/l #	85
20) Methylene Chloride	2.788	84	3051	2.861	ug/l #	90
25) 2-Butanone	4.574	43	6686	8.431	ug/l	96
43) Isopropyl Acetate	6.342	43	84809	29.769	ug/l	100
73) Isopropylbenzene	10.964	105	19117	3.722	ug/l	99
78) n-propylbenzene	11.305	91	43483	7.711	ug/l	100
95) Naphthalene	13.774	128	17356	3.637	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
Data File : VX043235.D
Acq On : 02 Oct 2024 16:41
Operator : JC/MD
Sample : P4232-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-R

Quant Time: Oct 03 04:48:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

