

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
 Data File : VX043236.D
 Acq On : 02 Oct 2024 17:04
 Operator : JC/MD
 Sample : P4236-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP4

Quant Time: Oct 03 04:54:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	81265	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	157243	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	135144	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56917	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72640	54.733	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.460%
35) Dibromofluoromethane	5.385	113	52371	48.294	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.580%
50) Toluene-d8	8.647	98	188998	50.292	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.580%
62) 4-Bromofluorobenzene	11.079	95	66514	48.719	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.440%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	30858	55.676	ug/l	100
18) Methyl Acetate	2.715	43	2018	1.451	ug/l	92
20) Methylene Chloride	2.788	84	3424	3.351	ug/l	97
25) 2-Butanone	4.586	43	4009	5.275	ug/l	95
43) Isopropyl Acetate	6.342	43	82433	30.336	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
Data File : VX043236.D
Acq On : 02 Oct 2024 17:04
Operator : JC/MD
Sample : P4236-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP4

Quant Time: Oct 03 04:54:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

