

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
 Data File : VX043237.D
 Acq On : 02 Oct 2024 17:27
 Operator : JC/MD
 Sample : P4236-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP3

Quant Time: Oct 03 05:00:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	85853	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	164952	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	147477	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64669	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	74890	53.412	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.820%
35) Dibromofluoromethane	5.385	113	54951	48.305	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.620%
50) Toluene-d8	8.647	98	200574	50.878	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.760%
62) 4-Bromofluorobenzene	11.079	95	73084	51.030	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.060%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	29353	50.130	ug/l	99
18) Methyl Acetate	2.715	43	2034	1.385	ug/l #	69
20) Methylene Chloride	2.788	84	3439	3.185	ug/l	93
25) 2-Butanone	4.587	43	4224	5.261	ug/l #	90
43) Isopropyl Acetate	6.342	43	80283	28.164	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
Data File : VX043237.D
Acq On : 02 Oct 2024 17:27
Operator : JC/MD
Sample : P4236-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP3

Quant Time: Oct 03 05:00:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

