

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
 Data File : VX043238.D
 Acq On : 02 Oct 2024 17:50
 Operator : JC/MD
 Sample : P4236-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP1

Quant Time: Oct 03 04:59:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	83361	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	159540	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	144115	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64304	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72891	53.541	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.080%			
35) Dibromofluoromethane	5.385	113	53028	48.196	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.400%			
50) Toluene-d8	8.647	98	193180	50.665	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.340%			
62) 4-Bromofluorobenzene	11.079	95	71536	51.643	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 103.280%			
Target Compounds						
16) Acetone	2.386	43	29714	52.264	ug/l	99
18) Methyl Acetate	2.709	43	2307	1.618	ug/l	94
20) Methylene Chloride	2.788	84	3313	3.160	ug/l	94
43) Isopropyl Acetate	6.342	43	81576	29.588	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100224\
Data File : VX043238.D
Acq On : 02 Oct 2024 17:50
Operator : JC/MD
Sample : P4236-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP1

Quant Time: Oct 03 04:59:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

