

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100225\
 Data File : VX047957.D
 Acq On : 02 Oct 2025 11:24
 Operator : JC/MD
 Sample : Q3246-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-GW-460-462

Quant Time: Oct 03 03:27:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	127914	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	256538	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	255834	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	122681	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	95997	47.148	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.300%		
35) Dibromofluoromethane	5.379	113	81995	47.658	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.320%		
50) Toluene-d8	8.635	98	282326	47.424	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	94.840%		
62) 4-Bromofluorobenzene	11.061	95	116514	51.570	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	103.140%		
Target Compounds						
16) Acetone	2.386	43	4455	5.032	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100225\
Data File : VX047957.D
Acq On : 02 Oct 2025 11:24
Operator : JC/MD
Sample : Q3246-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-GW-460-462

Quant Time: Oct 03 03:27:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

