

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100225\
 Data File : VX047961.D
 Acq On : 02 Oct 2025 12:46
 Operator : JC/MD
 Sample : Q3246-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-GW-500-502

Quant Time: Oct 03 03:29:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	143047	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.750	114	293439	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.036	117	309899	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	156960	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	124531	54.692	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.380%	
35) Dibromofluoromethane	5.373	113	98868	50.239	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.480%	
50) Toluene-d8	8.634	98	334661	49.146	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.300%	
62) 4-Bromofluorobenzene	11.067	95	146161	56.556	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 113.120%	
Target Compounds						
						Qvalue
3) Chloromethane	1.300	50	1310	0.673	ug/l	93
16) Acetone	2.379	43	10657	10.764	ug/l	95
25) 2-Butanone	4.562	43	2693	2.181	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100225\
Data File : VX047961.D
Acq On : 02 Oct 2025 12:46
Operator : JC/MD
Sample : Q3246-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-GW-500-502

Quant Time: Oct 03 03:29:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

