

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100225\
 Data File : VX047962.D
 Acq On : 02 Oct 2025 13:07
 Operator : JC/MD
 Sample : Q3246-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-DUP-20250925

Quant Time: Oct 03 03:29:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	135174	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	286335	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	299172	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	150587	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	117999	54.842	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.680%	
35) Dibromofluoromethane	5.379	113	95408	49.683	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.360%	
50) Toluene-d8	8.635	98	326742	49.173	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.340%	
62) 4-Bromofluorobenzene	11.061	95	142221	56.397	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 112.800%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.307	50	496	0.270	ug/l	95
16) Acetone	2.380	43	3821	4.084	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100225\
Data File : VX047962.D
Acq On : 02 Oct 2025 13:07
Operator : JC/MD
Sample : Q3246-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-DUP-20250925

Quant Time: Oct 03 03:29:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

