

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100225\
 Data File : VX047963.D
 Acq On : 02 Oct 2025 13:27
 Operator : JC/MD
 Sample : Q3246-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-GW-520-522

Quant Time: Oct 03 03:30:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	151967	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.750	114	317249	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.036	117	332386	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	167184	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	130160	53.809	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	107.620%	
35) Dibromofluoromethane	5.379	113	105359	49.519	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.040%	
50) Toluene-d8	8.634	98	358525	48.699	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	97.400%	
62) 4-Bromofluorobenzene	11.061	95	157385	56.329	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	112.660%	
Target Compounds						
16) Acetone	2.385	43	5114	4.862	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100225\
Data File : VX047963.D
Acq On : 02 Oct 2025 13:27
Operator : JC/MD
Sample : Q3246-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-GW-520-522

Quant Time: Oct 03 03:30:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

