

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100225\
 Data File : VX047964.D
 Acq On : 02 Oct 2025 13:48
 Operator : JC/MD
 Sample : Q3246-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-GW-540-542

Quant Time: Oct 03 03:30:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	131720	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	271856	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	280780	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	143261	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	111636	53.245	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.480%
35) Dibromofluoromethane	5.379	113	88993	48.811	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.620%
50) Toluene-d8	8.635	98	305278	48.390	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.780%
62) 4-Bromofluorobenzene	11.067	95	131671	54.994	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.980%
Target Compounds						
						Qvalue
3) Chloromethane	1.301	50	837	0.467	ug/l #	87
16) Acetone	2.380	43	6651	7.296	ug/l	93
19) Methyl tert-butyl Ether	3.111	73	21351	3.737	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100225\
Data File : VX047964.D
Acq On : 02 Oct 2025 13:48
Operator : JC/MD
Sample : Q3246-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-GW-540-542

Quant Time: Oct 03 03:30:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

