

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100225\
 Data File : VX047972.D
 Acq On : 02 Oct 2025 16:33
 Operator : JC/MD
 Sample : Q3194-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW11-MW01I-20250922

Quant Time: Oct 03 03:33:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	134837	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	279573	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	292400	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	149635	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	117345	54.674	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.340%
35) Dibromofluoromethane	5.373	113	94906	50.617	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.240%
50) Toluene-d8	8.635	98	316318	48.756	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.520%
62) 4-Bromofluorobenzene	11.067	95	140496	57.061	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.120%
Target Compounds						
16) Acetone	2.374	43	2674	2.865	ug/l	Qvalue # 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100225\
Data File : VX047972.D
Acq On : 02 Oct 2025 16:33
Operator : JC/MD
Sample : Q3194-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW11-MW01I-20250922

Quant Time: Oct 03 03:33:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

