

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100225\
 Data File : VX047976.D
 Acq On : 02 Oct 2025 17:56
 Operator : JC/MD
 Sample : Q3194-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW11-MW01S-20250922

Quant Time: Oct 03 03:34:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	119521	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	250300	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	260092	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	131301	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	105665	55.541	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.080%
35) Dibromofluoromethane	5.379	113	84455	50.311	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.620%
50) Toluene-d8	8.634	98	283891	48.875	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.760%
62) 4-Bromofluorobenzene	11.067	95	120098	54.481	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.960%
Target Compounds						
44) Trichloroethene	7.116	130	453	0.252	ug/l	77
52) Toluene	8.708	92	1281	0.279	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100225\
Data File : VX047976.D
Acq On : 02 Oct 2025 17:56
Operator : JC/MD
Sample : Q3194-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW11-MW01S-20250922

Quant Time: Oct 03 03:34:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

