

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100318\
 Data File : VX004958.D
 Acq On : 03 Oct 2018 10:13
 Operator : JC/MD
 Sample : J5084-06 10X
 Misc : 5.91g/5mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-10(7-7.5)

Manual Integrations
 APPROVED

sam
 10/3/2018 5:57:05 PM

Quant Time: Oct 03 16:45:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	259351	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	416435	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	411547	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	246356	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	194493	52.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.00%	
35) Dibromofluoromethane	5.51	113	161426	45.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.48%	
50) Toluene-d8	8.72	98	599028	51.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.10%	
62) 4-Bromofluorobenzene	11.14	95	232192	54.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.86%	
Target Compounds						
52) Toluene	8.79	92	29609	3.574	ug/l	96
67) Ethyl Benzene	10.25	91	20096	1.199	ug/l	94
68) m/p-Xylenes	10.36	106	33419	5.104	ug/l	97
69) o-Xylene	10.70	106	11612	1.931	ug/l	97
80) 1,3,5-Trimethylbenzene	11.51	105	18154	1.082	ug/l	99
84) 1,2,4-Trimethylbenzene	11.81	105	61612	4.016	ug/l	98
95) Naphthalene	13.83	128	21734	1.765	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100318\
Data File : VX004958.D
Acq On : 03 Oct 2018 10:13
Operator : JC/MD
Sample : J5084-06 10X
Misc : 5.91µ/5mL/100µL/5.0mL/MSVOA_X/MEOH
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-10(7-7.5)

Manual Integrations
APPROVED

sam
10/3/2018 5:57:05 PM

Quant Time: Oct 03 16:45:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

