

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100318\
 Data File : VX004961.D
 Acq On : 03 Oct 2018 11:33
 Operator : JC/MD
 Sample : J5084-09 50X
 Misc : 5.31g/5mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-9(6-6.5)

Quant Time: Oct 03 15:17:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	272499	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	439672	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	416439	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	236409	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	188539	48.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.86%	
35) Dibromofluoromethane	5.50	113	161414	43.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.64%	
50) Toluene-d8	8.71	98	631209	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
62) 4-Bromofluorobenzene	11.14	95	242380	54.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.62%	
Target Compounds						
					Qvalue	
39) Methylcyclohexane	7.46	83	65965	12.707	ug/l	92
52) Toluene	8.79	92	36331	4.154	ug/l	95
67) Ethyl Benzene	10.26	91	374804	22.107	ug/l	95
68) m/p-Xylenes	10.36	106	700889	105.794	ug/l	96
69) o-Xylene	10.70	106	284062	46.682	ug/l	99
73) Isopropylbenzene	11.02	105	104447	7.009	ug/l	99
78) n-propylbenzene	11.36	91	510725	29.872	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	1148574	84.234	ug/l	100
84) 1,2,4-Trimethylbenzene	11.81	105	3508796	249.914	ug/l	99
85) sec-Butylbenzene	11.94	105	105544	7.063	ug/l	93
86) p-Isopropyltoluene	12.07	119	65423	4.832	ug/l	99
89) n-Butylbenzene	12.38	91	453954	36.919	ug/l #	26
95) Naphthalene	13.83	128	482964	24.926	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100318\
Data File : VX004961.D
Acq On : 03 Oct 2018 11:33
Operator : JC/MD
Sample : J5084-09 50X
Misc : 5.31q/5mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-9(6-6.5)

Quant Time: Oct 03 15:17:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

