


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX100318\
Data File : VX004921.D
Acq On : 02 Oct 2018 16:10
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Title : SW846 8260
Last Update : Tue Oct 02 02:04:34 2018



AutoFind: Scans 1637, 1638, 1639; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.9	8151	PASS
75	95	30	60	48.7	22168	PASS
95	95	100	100	100.0	45485	PASS
96	95	5	9	6.8	3097	PASS
173	174	0.00	2	1.3	546	PASS
174	95	50	100	93.5	42538	PASS
175	174	5	9	7.3	3120	PASS
176	174	95	101	97.9	41648	PASS
177	176	5	9	6.7	2781	PASS