


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX100318\
Data File : VX004948.D
Acq On : 03 Oct 2018 04:50
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Title : SW846 8260
Last Update : Tue Oct 02 02:04:34 2018



AutoFind: Scans 1637, 1638, 1639; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.9	7476	PASS
75	95	30	60	48.0	18956	PASS
95	95	100	100	100.0	39530	PASS
96	95	5	9	6.8	2676	PASS
173	174	0.00	2	1.6	588	PASS
174	95	50	100	94.7	37416	PASS
175	174	5	9	7.6	2857	PASS
176	174	95	101	95.4	35698	PASS
177	176	5	9	6.9	2446	PASS