

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100319\
 Data File : VX012779.D
 Acq On : 03 Oct 2019 13:30
 Operator : JC/SP
 Sample : K5157-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-TB-2019-7

Quant Time: Oct 04 07:03:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	180617	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	310247	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	267748	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	98918	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	126122	47.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.90%	
35) Dibromofluoromethane	5.49	113	98273	53.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.38%	
50) Toluene-d8	8.71	98	373264	52.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.92%	
62) 4-Bromofluorobenzene	11.14	95	123033	43.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.94%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100319\
Data File : VX012779.D
Acq On : 03 Oct 2019 13:30
Operator : JC/SP
Sample : K5157-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-TB-2019-7

Quant Time: Oct 04 07:03:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

