

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100319\
 Data File : VX012782.D
 Acq On : 03 Oct 2019 14:40
 Operator : JC/SP
 Sample : K5157-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0524

Quant Time: Oct 04 07:11:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	174286	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	299541	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	250857	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	91478	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	121394	47.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.66%	
35) Dibromofluoromethane	5.48	113	91859	51.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.98%	
50) Toluene-d8	8.71	98	353003	51.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.76%	
62) 4-Bromofluorobenzene	11.14	95	115545	42.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.54%	
Target Compounds						
4) Vinyl Chloride	1.40	62	2684	1.199	ug/l	96
21) trans-1,2-Dichloroethene	3.17	96	2543	1.225	ug/l #	77
24) 1,1-Dichloroethane	3.69	63	10234	2.621	ug/l	96
27) cis-1,2-Dichloroethene	4.59	96	9523	3.953	ug/l	86
44) Trichloroethene	7.21	130	1212	0.558	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100319\
Data File : VX012782.D
Acq On : 03 Oct 2019 14:40
Operator : JC/SP
Sample : K5157-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-0524

Quant Time: Oct 04 07:11:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

