

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100322\
 Data File : VX031845.D
 Acq On : 03 Oct 2022 19:48
 Operator : JC/MD
 Sample : N4860-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2R

Quant Time: Oct 04 07:11:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 06:59:30 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	77489	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	121997	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72222	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75949	54.708	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.420%	
35) Dibromofluoromethane	5.385	113	69144	47.703	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.400%	
50) Toluene-d8	8.653	98	239557	46.200	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 92.400%	
62) 4-Bromofluorobenzene	11.079	95	80219	42.962	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 85.920%	
Target Compounds						
						Qvalue
73) Isopropylbenzene	10.963	105	3828	0.750	ug/l	100
83) tert-Butylbenzene	11.713	119	1892	0.439	ug/l	87
85) sec-Butylbenzene	11.890	105	4449	0.858	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100322\
Data File : VX031845.D
Acq On : 03 Oct 2022 19:48
Operator : JC/MD
Sample : N4860-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2R

Quant Time: Oct 04 07:11:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 06:59:30 2022
Response via : Initial Calibration

