

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX100322\  
 Data File : VX031846.D  
 Acq On : 03 Oct 2022 20:12  
 Operator : JC/MD  
 Sample : N4864-08  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 24 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 MW-13

Quant Time: Oct 04 07:12:15 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X100322W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Oct 04 06:59:30 2022  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 5.550  | 168  | 85072    | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 6.763  | 114  | 126994   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 10.055 | 117  | 137449   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.024 | 152  | 92995    | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |         |
|-----------------------------|--------|-------|----------|----------|------|---------|
| System Monitoring Compounds |        |       |          |          |      |         |
| 33) 1,2-Dichloroethane-d4   | 5.952  | 65    | 75774    | 49.717   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =    | 99.440% |
| 35) Dibromofluoromethane    | 5.385  | 113   | 70970    | 47.037   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =    | 94.080% |
| 50) Toluene-d8              | 8.647  | 98    | 260634   | 48.287   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =    | 96.580% |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 89773    | 46.186   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =    | 92.380% |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX100322\  
Data File : VX031846.D  
Acq On : 03 Oct 2022 20:12  
Operator : JC/MD  
Sample : N4864-08  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 24 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-13

Quant Time: Oct 04 07:12:15 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X100322W.M  
Quant Title : SW846 8260  
QLast Update : Tue Oct 04 06:59:30 2022  
Response via : Initial Calibration

