

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100322\
 Data File : VX031847.D
 Acq On : 03 Oct 2022 20:35
 Operator : JC/MD
 Sample : N4860-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-04

Quant Time: Oct 04 07:12:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 06:59:30 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	79974	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	123256	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	132968	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86664	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	75324	52.572	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.140%
35) Dibromofluoromethane	5.385	113	69103	47.188	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.380%
50) Toluene-d8	8.647	98	250213	47.762	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.520%
62) 4-Bromofluorobenzene	11.079	95	85178	45.151	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.300%
Target Compounds						
19) Methyl tert-butyl Ether	3.111	73	3055	0.840	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100322\
Data File : VX031847.D
Acq On : 03 Oct 2022 20:35
Operator : JC/MD
Sample : N4860-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-04

Quant Time: Oct 04 07:12:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 06:59:30 2022
Response via : Initial Calibration

