

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100322\
 Data File : VX031851.D
 Acq On : 03 Oct 2022 22:07
 Operator : JC/MD
 Sample : PB148021TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB148021TB

Quant Time: Oct 04 08:02:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 06:59:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	87209	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	135065	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	159541	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102228	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	79784	51.065	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.140%
35) Dibromofluoromethane	5.385	113	74487	46.417	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.840%
50) Toluene-d8	8.647	98	281886	49.104	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.200%
62) 4-Bromofluorobenzene	11.079	95	99284	48.027	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.060%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	8768	7.362	ug/l	94
20) Methylene Chloride	2.788	84	7406	5.418	ug/l #	84
37) Ethyl Acetate	4.714	43	14998	9.958	ug/l #	97
43) Isopropyl Acetate	6.342	43	66300	27.225	ug/l #	91
80) 1,3,5-Trimethylbenzene	11.457	105	1809	0.300	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100322\
Data File : VX031851.D
Acq On : 03 Oct 2022 22:07
Operator : JC/MD
Sample : PB148021TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB148021TB

Quant Time: Oct 04 08:02:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 06:59:30 2022
Response via : Initial Calibration

