

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100322\
 Data File : VX031852.D
 Acq On : 03 Oct 2022 22:30
 Operator : JC/MD
 Sample : PB148021ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB148021ZHE#01

Quant Time: Oct 04 08:02:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 06:59:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	81154	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	128851	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	149558	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95818	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77511	53.312	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.620%
35) Dibromofluoromethane	5.385	113	72627	47.441	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.880%
50) Toluene-d8	8.653	98	266066	48.583	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.160%
62) 4-Bromofluorobenzene	11.079	95	94248	47.790	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.580%
Target Compounds						
37) Ethyl Acetate	4.715	43	14390	10.015	ug/l #	94
43) Isopropyl Acetate	6.342	43	66206	28.497	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100322\
Data File : VX031852.D
Acq On : 03 Oct 2022 22:30
Operator : JC/MD
Sample : PB148021ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB148021ZHE#01

Quant Time: Oct 04 08:02:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 06:59:30 2022
Response via : Initial Calibration

