

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100322\
 Data File : VX031853.D
 Acq On : 03 Oct 2022 22:54
 Operator : JC/MD
 Sample : PB148021ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB148021ZHE#02

Quant Time: Oct 04 08:02:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 06:59:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	82094	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	128439	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153867	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97115	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	77903	52.968	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.940%
35) Dibromofluoromethane	5.391	113	71386	46.780	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.560%
50) Toluene-d8	8.646	98	276562	50.662	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.320%
62) 4-Bromofluorobenzene	11.079	95	95376	48.517	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.040%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	9880	12.485	ug/l #	84
20) Methylene Chloride	2.788	84	8303	6.453	ug/l #	83
37) Ethyl Acetate	4.720	43	14688	10.255	ug/l #	94
43) Isopropyl Acetate	6.342	43	67149	28.996	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100322\
Data File : VX031853.D
Acq On : 03 Oct 2022 22:54
Operator : JC/MD
Sample : PB148021ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB148021ZHE#02

Quant Time: Oct 04 08:02:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 06:59:30 2022
Response via : Initial Calibration

