

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100322\
 Data File : VX031856.D
 Acq On : 04 Oct 2022 00:03
 Operator : JC/MD
 Sample : N4932-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RBR-2520555

Quant Time: Oct 04 08:04:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 06:59:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	101322	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	152847	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	172202	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	114743	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84115	46.338	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	92.680%		
35) Dibromofluoromethane	5.385	113	83405	45.928	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.860%		
50) Toluene-d8	8.647	98	304669	46.898	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	93.800%		
62) 4-Bromofluorobenzene	11.079	95	107987	46.160	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.320%		
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	9958	6.777	ug/l #	87
20) Methylene Chloride	2.788	84	9412	5.927	ug/l #	78
37) Ethyl Acetate	4.727	43	12696	7.449	ug/l #	93
43) Isopropyl Acetate	6.342	43	57859	20.994	ug/l #	90
95) Naphthalene	13.774	128	55435	7.198	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100322\
Data File : VX031856.D
Acq On : 04 Oct 2022 00:03
Operator : JC/MD
Sample : N4932-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RBR-2520555

Quant Time: Oct 04 08:04:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 06:59:30 2022
Response via : Initial Calibration

