

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100322\
 Data File : VX031859.D
 Acq On : 04 Oct 2022 01:12
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Oct 04 08:06:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 06:59:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	85570	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	136178	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	142052	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89454	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87148	56.847	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.700%
35) Dibromofluoromethane	5.385	113	85445	52.811	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.620%
50) Toluene-d8	8.647	98	254513	43.973	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	87.940%
62) 4-Bromofluorobenzene	11.079	95	88009	42.225	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.460%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100322\
Data File : VX031859.D
Acq On : 04 Oct 2022 01:12
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Oct 04 08:06:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 06:59:30 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX031859.D
Acq: 04 Oct 2022 01:12




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:168 Resp: 85570
Ion Ratio Lower Upper
168 100
99 49.8 48.8 73.2



#33
1,2-Dichloroethane-d4
Concen: 56.847 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.006 min
Lab File: VX031859.D
Acq: 04 Oct 2022 01:12

Tgt Ion: 65 Resp: 87148
Ion Ratio Lower Upper
65 100
67 51.9 0.0 102.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 911
Delta R.T. 0.000 min
Lab File: VX031859.D
Acq: 04 Oct 2022 01:12




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:114 Resp: 136178
Ion Ratio Lower Upper
114 100
63 15.5 0.0 42.6
88 14.3 0.0 33.0



#35
Dibromofluoromethane
Concen: 52.811 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX031859.D
Acq: 04 Oct 2022 01:12

Tgt Ion:113 Resp: 85445
Ion Ratio Lower Upper
113 100
111 102.7 82.4 123.6
192 21.8 14.1 21.1#









#50
Toluene-d8
Concen: 43.973 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX031859.D
Acq: 04 Oct 2022 01:12




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 98 Resp: 254513
Ion Ratio Lower Upper
98 100
100 67.0 52.5 78.7



#62
4-Bromofluorobenzene
Concen: 42.225 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX031859.D
Acq: 04 Oct 2022 01:12

Tgt Ion: 95 Resp: 88009
Ion Ratio Lower Upper
95 100
174 94.1 0.0 153.2
176 89.9 0.0 144.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX031859.D
Acq: 04 Oct 2022 01:12

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Tgt Ion:117 Resp: 142052
Ion Ratio Lower Upper
117 100
82 47.9 46.8 70.2
119 31.8 26.4 39.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX031859.D
Acq: 04 Oct 2022 01:12

Tgt Ion:152 Resp: 89454
Ion Ratio Lower Upper
152 100
115 55.6 43.4 130.1
150 155.6 0.0 346.8

