

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
 Data File : VX043270.D
 Acq On : 03 Oct 2024 18:17
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 04 01:36:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	112786	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	187556	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	158082	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	75329	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	78054	42.375	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.760%
35) Dibromofluoromethane	5.385	113	58094	44.914	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.820%
50) Toluene-d8	8.647	98	193679	43.208	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	86.420%
62) 4-Bromofluorobenzene	11.079	95	70385	43.222	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	86.440%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	60369	46.016	ug/l	98
3) Chloromethane	1.294	50	56954	42.152	ug/l	100
4) Vinyl Chloride	1.374	62	61790	45.656	ug/l	97
5) Bromomethane	1.599	94	22062	39.901	ug/l	100
6) Chloroethane	1.666	64	21974	43.600	ug/l	99
7) Trichlorofluoromethane	1.867	101	92110	45.551	ug/l	99
8) Diethyl Ether	2.136	74	36977	46.990	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.319	101	59339	47.024	ug/l	99
10) Methyl Iodide	2.447	142	69372	43.769	ug/l	100
11) Tert butyl alcohol	3.001	59	86234	246.020	ug/l	100
12) 1,1-Dichloroethene	2.312	96	56193	44.843	ug/l	99
13) Acrolein	2.239	56	76877	232.213	ug/l	99
14) Allyl chloride	2.660	41	92598	44.978	ug/l	99
15) Acrylonitrile	3.068	53	169381	236.653	ug/l	99
16) Acetone	2.392	43	157984	205.382	ug/l	99
17) Carbon Disulfide	2.501	76	123107	40.390	ug/l	97
18) Methyl Acetate	2.709	43	97172	50.357	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	223641	49.884	ug/l	99
20) Methylene Chloride	2.788	84	63225	44.579	ug/l	96
21) trans-1,2-Dichloroethene	3.087	96	59191	47.462	ug/l	97
22) Diisopropyl ether	3.763	45	200291	48.370	ug/l	97
23) Vinyl Acetate	3.721	43	1209338	252.310	ug/l	99
24) 1,1-Dichloroethane	3.605	63	115920	48.053	ug/l	99
25) 2-Butanone	4.562	43	242939	230.338	ug/l	100
26) 2,2-Dichloropropane	4.471	77	103677	45.171	ug/l	100
27) cis-1,2-Dichloroethene	4.489	96	71759	46.137	ug/l	98
28) Bromochloromethane	4.903	49	39613	41.623	ug/l	100
29) Tetrahydrofuran	5.013	42	156932	233.566	ug/l	100
30) Chloroform	5.092	83	124843	47.125	ug/l	100
31) Cyclohexane	5.464	56	91922	46.249	ug/l	99
32) 1,1,1-Trichloroethane	5.379	97	116717	47.967	ug/l	100
36) 1,1-Dichloropropene	5.690	75	79424	48.605	ug/l	99
37) Ethyl Acetate	4.721	43	98089	49.987	ug/l	99
38) Carbon Tetrachloride	5.672	117	97655	50.483	ug/l	99
39) Methylcyclohexane	7.379	83	100069	48.692	ug/l	99
40) Benzene	6.031	78	242940	49.037	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
 Data File : VX043270.D
 Acq On : 03 Oct 2024 18:17
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 04 01:36:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	51787	50.698	ug/l	96
42) 1,2-Dichloroethane	6.086	62	102384	50.714	ug/l	100
43) Isopropyl Acetate	6.342	43	160966	49.663	ug/l	99
44) Trichloroethene	7.123	130	65224	49.636	ug/l	94
45) 1,2-Dichloropropane	7.427	63	60143	49.467	ug/l	99
46) Dibromomethane	7.580	93	48615	49.900	ug/l	98
47) Bromodichloromethane	7.818	83	96947	50.614	ug/l	98
48) Methyl methacrylate	7.696	41	78328	49.253	ug/l	99
49) 1,4-Dioxane	7.665	88	34065	1101.934	ug/l	95
51) 4-Methyl-2-Pentanone	8.573	43	488840	248.328	ug/l	99
52) Toluene	8.714	92	150463	48.458	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	94369	49.581	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	100271	49.119	ug/l	98
55) 1,1,2-Trichloroethane	9.153	97	60932	49.378	ug/l	97
56) Ethyl methacrylate	9.116	69	103028	51.068	ug/l	99
57) 1,3-Dichloropropane	9.305	76	103540	49.264	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	232712	255.903	ug/l	100
59) 2-Hexanone	9.433	43	372722	250.379	ug/l	99
60) Dibromochloromethane	9.518	129	69090	48.262	ug/l	98
61) 1,2-Dibromoethane	9.610	107	64209	48.430	ug/l	100
64) Tetrachloroethene	9.275	164	53321	48.611	ug/l	96
65) Chlorobenzene	10.079	112	163490	48.105	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.165	131	60221	50.111	ug/l	99
67) Ethyl Benzene	10.195	91	287881	49.007	ug/l	99
68) m/p-Xylenes	10.299	106	214462	96.606	ug/l	98
69) o-Xylene	10.640	106	106831	48.428	ug/l	99
70) Styrene	10.652	104	180408	50.463	ug/l	99
71) Bromoform	10.799	173	46942	50.991	ug/l #	99
73) Isopropylbenzene	10.963	105	279425	47.711	ug/l	99
74) N-amyl acetate	10.841	43	136527	49.107	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.213	83	93555	49.035	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	106744	63.885	ug/l	82
77) Bromobenzene	11.195	156	67752	47.537	ug/l	100
78) n-propylbenzene	11.305	91	315475	49.068	ug/l	100
79) 2-Chlorotoluene	11.366	91	194540	47.065	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	235295	48.704	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	29193	45.280	ug/l	95
82) 4-Chlorotoluene	11.451	91	226491	48.101	ug/l	100
83) tert-Butylbenzene	11.713	119	238251	47.868	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	233919	48.122	ug/l	97
85) sec-Butylbenzene	11.890	105	287933	48.800	ug/l	99
86) p-Isopropyltoluene	12.006	119	239521	48.657	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	121233	48.287	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	120562	46.577	ug/l	99
89) n-Butylbenzene	12.329	91	206130	50.016	ug/l	99
90) Hexachloroethane	12.536	117	44782	48.507	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	122452	47.691	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	22322	47.920	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	73386	49.578	ug/l	100
94) Hexachlorobutadiene	13.725	225	29692	50.601	ug/l	98
95) Naphthalene	13.774	128	262695	48.276	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	73595	48.189	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
Data File : VX043270.D
Acq On : 03 Oct 2024 18:17
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Oct 04 01:36:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 04 01:36:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :

MSVOA_X

ClientSampleId :







Tgt Ion:168 Resp: 112786
Ion Ratio Lower Upper
168 100
99 65.3 49.7 74.5



#2
Dichlorodifluoromethane
Concen: 46.016 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 85 Resp: 60369
Ion Ratio Lower Upper
85 100
87 32.7 17.1 51.2








#3
Chloromethane
Concen: 42.152 ug/l
RT: 1.294 min Scan# 34
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 50 Resp: 56954
Ion Ratio Lower Upper
50 100
52 32.8 26.1 39.1



#4
Vinyl Chloride
Concen: 45.656 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 62 Resp: 61790
Ion Ratio Lower Upper
62 100
64 31.6 26.8 40.2









#5
 Bromomethane
 Concen: 39.901 ug/l
 RT: 1.599 min Scan# 84
 Delta R.T. 0.000 min
 Lab File: VX043270.D
 Acq: 03 Oct 2024 18:17




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 94 Resp: 22062
 Ion Ratio Lower Upper
 94 100
 96 96.0 76.6 115.0



#6
 Chloroethane
 Concen: 43.600 ug/l
 RT: 1.666 min Scan# 95
 Delta R.T. -0.000 min
 Lab File: VX043270.D
 Acq: 03 Oct 2024 18:17

Tgt Ion: 64 Resp: 21974
 Ion Ratio Lower Upper
 64 100
 66 33.1 25.8 38.8









#7
Trichlorofluoromethane
Concen: 45.551 ug/l
RT: 1.867 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 92110
Ion Ratio Lower Upper
101 100
103 65.2 52.7 79.1



#8
Diethyl Ether
Concen: 46.990 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 74 Resp: 36977
Ion Ratio Lower Upper
74 100
45 97.6 47.1 141.4









#9
1,1,2-Trichlorotrifluoroethane
Concen: 47.024 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 59339
Ion Ratio Lower Upper
101 100
85 46.0 37.0 55.6
151 73.4 60.0 90.0



#10
Methyl Iodide
Concen: 43.769 ug/l
RT: 2.447 min Scan# 223
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion:142 Resp: 69372
Ion Ratio Lower Upper
142 100
127 50.2 39.8 59.8
141 14.6 11.6 17.4





#11

Tert butyl alcohol

Concen: 246.020 ug/l

RT: 3.001 min Scan# 313

Delta R.T. 0.006 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 59 Resp: 86234

Ion Ratio Lower Upper

59 100

57 10.8 8.6 13.0



#12

1,1-Dichloroethene

Concen: 44.843 ug/l

RT: 2.312 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17




Tgt Ion: 96 Resp: 56193


Ion Ratio Lower Upper

96 100

61 164.1 130.6 195.8

98 66.0 51.7 77.5





#13

Acrolein

Concen: 232.213 ug/l

RT: 2.239 min Scan# 189

Delta R.T. -0.000 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 56 Resp: 76877

Ion Ratio Lower Upper

56 100

55 70.9 57.4 86.0



#14

Allyl chloride

Concen: 44.978 ug/l

RT: 2.660 min Scan# 258

Delta R.T. -0.000 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17




Tgt Ion: 41 Resp: 92598

Ion Ratio Lower Upper

41 100

39 76.7 61.7 92.5

76 34.5 28.6 43.0









#15
Acrylonitrile
Concen: 236.653 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 53 Resp: 169381
Ion Ratio Lower Upper
53 100
52 83.8 67.6 101.4
51 37.3 30.3 45.5



#16
Acetone
Concen: 205.382 ug/l
RT: 2.392 min Scan# 214
Delta R.T. 0.006 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 43 Resp: 157984
Ion Ratio Lower Upper
43 100
58 28.8 22.6 33.8









#17
Carbon Disulfide
Concen: 40.390 ug/l
RT: 2.501 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 123107
Ion Ratio Lower Upper
76 100
78 9.5 6.7 10.1



#18
Methyl Acetate
Concen: 50.357 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 43 Resp: 97172
Ion Ratio Lower Upper
43 100
74 23.6 19.1 28.7









#19
Methyl tert-butyl Ether
Concen: 49.884 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 73 Resp: 223641
Ion Ratio Lower Upper
73 100
57 21.4 16.7 25.1



#20
Methylene Chloride
Concen: 44.579 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 84 Resp: 63225
Ion Ratio Lower Upper
84 100
49 118.9 96.7 145.1
51 35.9 30.4 45.6
86 62.5 54.7 82.1









#21
trans-1,2-Dichloroethene
Concen: 47.462 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:	96	Resp:	59191
Ion	Ratio	Lower	Upper
96	100		
61	140.9	112.3	168.5
98	61.4	54.6	82.0



#22
Diisopropyl ether
Concen: 48.370 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.001 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion:	45	Resp:	200291
Ion	Ratio	Lower	Upper
45	100		
43	89.3	68.5	102.7
87	28.1	23.0	34.4
59	11.7	8.8	13.2









#23
 Vinyl Acetate
 Concen: 252.310 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX043270.D
 Acq: 03 Oct 2024 18:17




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 1209338
 Ion Ratio Lower Upper
 43 100
 86 10.5 8.8 13.2



#24
 1,1-Dichloroethane
 Concen: 48.053 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX043270.D
 Acq: 03 Oct 2024 18:17

Tgt Ion: 63 Resp: 115920
 Ion Ratio Lower Upper
 63 100
 98 7.1 3.8 11.2
 100 4.9 2.5 7.4








#25
2-Butanone
Concen: 230.338 ug/l
RT: 4.562 min Scan# 570
Delta R.T. 0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 242939
Ion Ratio Lower Upper
43 100
72 25.6 20.6 31.0



#26
2,2-Dichloropropane
Concen: 45.171 ug/l
RT: 4.471 min Scan# 555
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 77 Resp: 103677
Ion Ratio Lower Upper
77 100
97 21.6 10.8 32.3





#27

cis-1,2-Dichloroethene

Concen: 46.137 ug/l

RT: 4.489 min Scan# 557

Delta R.T. 0.006 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 96 Resp: 71759

Ion Ratio Lower Upper

96 100

61 151.9 0.0 298.2

98 65.9 0.0 128.6



#28

Bromochloromethane

Concen: 41.623 ug/l

RT: 4.903 min Scan# 626

Delta R.T. 0.006 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17




Tgt Ion: 49 Resp: 39613


Ion Ratio Lower Upper

49 100

129 2.0 0.0 4.0

130 82.3 65.5 98.3





#29

Tetrahydrofuran

Concen: 233.566 ug/l

RT: 5.013 min Scan# 64

Delta R.T. 0.006 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 42 Resp: 156932

Ion Ratio Lower Upper

42 100

72 46.8 37.7 56.5

71 43.3 34.6 52.0



#30

Chloroform

Concen: 47.125 ug/l

RT: 5.092 min Scan# 657

Delta R.T. -0.000 min

Lab File: VX043270.D




Acq: 03 Oct 2024 18:17


Tgt Ion: 83 Resp: 124843

Ion Ratio Lower Upper

83 100


85 65.4 52.5 78.7








#31
Cyclohexane
Concen: 46.249 ug/l
RT: 5.464 min Scan# 718
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 56 Resp: 91922




Ion	Ratio	Lower	Upper
56	100		
69	30.8	25.8	38.6
84	88.3	71.1	106.7



#32
1,1,1-Trichloroethane
Concen: 47.967 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 97 Resp: 116717

Ion	Ratio	Lower	Upper
97	100		
99	63.7	51.2	76.8
61	44.8	35.6	53.4





#33

1,2-Dichloroethane-d4

Concen: 42.375 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 78054

Ion Ratio Lower Upper

65 100

67 49.8 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17




Tgt Ion: 114 Resp: 187556

Ion Ratio Lower Upper

114 100

63 21.1 0.0 43.0

88 17.0 0.0 33.0









#35
Dibromofluoromethane
Concen: 44.914 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:113 Resp: 58094
Ion Ratio Lower Upper
113 100
111 100.1 83.4 125.0
192 17.6 14.1 21.1



#36
1,1-Dichloropropene
Concen: 48.605 ug/l
RT: 5.690 min Scan# 755
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 75 Resp: 79424
Ion Ratio Lower Upper
75 100
110 34.8 17.8 53.3
77 31.2 24.9 37.3









#37
Ethyl Acetate
Concen: 49.987 ug/l
RT: 4.721 min Scan# 596
Delta R.T. 0.006 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 98089
Ion Ratio Lower Upper
43 100
61 13.3 11.0 16.4
70 10.8 8.9 13.3



#38
Carbon Tetrachloride
Concen: 50.483 ug/l
RT: 5.672 min Scan# 752
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17


Tgt Ion: 117 Resp: 97655
Ion Ratio Lower Upper
117 100
119 96.0 77.7 116.5
121 30.2 24.3 36.5








#39
Methylcyclohexane
Concen: 48.692 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. 0.006 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 83 Resp: 100069
Ion Ratio Lower Upper
83 100
55 74.9 59.4 89.2
98 49.6 39.3 58.9



#40
Benzene
Concen: 49.037 ug/l
RT: 6.031 min Scan# 811
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 78 Resp: 242940
Ion Ratio Lower Upper
78 100
77 23.9 18.8 28.2









#41
Methacrylonitrile
Concen: 50.698 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 41	Resp: 51787
Ion Ratio	Lower Upper
41 100	
39 61.1	45.0 67.4
67 72.8	56.0 84.0
52 30.9	25.6 38.4



#42
1,2-Dichloroethane
Concen: 50.714 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 62	Resp: 102384
Ion Ratio	Lower Upper
62 100	
98 7.9	0.0 16.0








#43
Isopropyl Acetate
Concen: 49.663 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 160966
Ion Ratio Lower Upper
43 100
61 18.6 15.1 22.7
87 13.3 10.8 16.2



#44
Trichloroethene
Concen: 49.636 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion:130 Resp: 65224
Ion Ratio Lower Upper
130 100
95 101.7 0.0 192.2





#45

1,2-Dichloropropane

Concen: 49.467 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 63 Resp: 60143

Ion Ratio Lower Upper

63 100

65 29.5 23.2 34.8



#46

Dibromomethane

Concen: 49.900 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.000 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17




Tgt Ion: 93 Resp: 48615

Ion Ratio Lower Upper

93 100

95 83.9 65.8 98.8

174 95.2 74.5 111.7









#47
Bromodichloromethane
Concen: 50.614 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 96947
Ion Ratio Lower Upper
83 100
85 62.8 51.4 77.2
127 7.5 6.2 9.4



#48
Methyl methacrylate
Concen: 49.253 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 41 Resp: 78328
Ion Ratio Lower Upper
41 100
69 79.6 64.1 96.1
39 66.4 52.8 79.2





#49

1,4-Dioxane

Concen: 1101.934 ug/l

RT: 7.665 min Scan# 1101

Delta R.T. -0.000 min


Lab File: VX043270.D

Acq: 03 Oct 2024 18:17




Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion:	88	Resp:	34065
Ion	Ratio	Lower	Upper
88	100		
43	34.8	29.5	44.3
58	64.4	55.2	82.8



#50

Toluene-d8

Concen: 43.208 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. -0.000 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17

Tgt Ion:	98	Resp:	193679
Ion	Ratio	Lower	Upper
98	100		
100	65.3	52.4	78.6









#51
4-Methyl-2-Pentanone
Concen: 248.328 ug/l
RT: 8.573 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 488840
Ion Ratio Lower Upper
43 100
58 38.2 30.0 45.0



#52
Toluene
Concen: 48.458 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 92 Resp: 150463
Ion Ratio Lower Upper
92 100
91 172.0 137.0 205.4





#53

t-1,3-Dichloropropene

Concen: 49.581 ug/l

RT: 8.976 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 75 Resp: 94369

Ion Ratio Lower Upper

75 100

77 30.3 25.1 37.7



#54

cis-1,3-Dichloropropene

Concen: 49.119 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17




Tgt Ion: 75 Resp: 100271


Ion Ratio Lower Upper

75 100

77 30.3 26.0 39.0


39 56.3 45.8 68.8








#55
1,1,2-Trichloroethane
Concen: 49.378 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:	97	Resp:	60932
Ion	Ratio	Lower	Upper
97	100		
83	80.3	66.9	100.3
85	52.7	43.9	65.9
99	61.9	51.6	77.4




#56
Ethyl methacrylate
Concen: 51.068 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17



Tgt Ion:	69	Resp:	103028
Ion	Ratio	Lower	Upper
69	100		
41	71.2	56.1	84.1
39	45.0	36.4	54.6





#57

1,3-Dichloropropane

Concen: 49.264 ug/l

RT: 9.305 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 103540

Ion Ratio Lower Upper

76 100

78 32.4 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 255.903 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX043270.D




Acq: 03 Oct 2024 18:17


Tgt Ion: 63 Resp: 232712

Ion Ratio Lower Upper

63 100

106 29.8 23.7 35.5





#59

2-Hexanone

Concen: 250.379 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. -0.000 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 372722

Ion Ratio Lower Upper

43 100

58 52.6 26.1 78.3



#60

Dibromochloromethane

Concen: 48.262 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX043270.D




Acq: 03 Oct 2024 18:17


Tgt Ion:129 Resp: 69090

Ion Ratio Lower Upper

129 100

127 78.1 38.3 114.9








#61
1,2-Dibromoethane
Concen: 48.430 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 64209
Ion Ratio Lower Upper
107 100
109 94.8 76.1 114.1



#62
4-Bromofluorobenzene
Concen: 43.222 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 95 Resp: 70385
Ion Ratio Lower Upper
95 100
174 74.7 0.0 145.0
176 72.1 0.0 143.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:117 Resp: 158082
Ion Ratio Lower Upper
117 100
82 53.3 45.0 67.6
119 31.2 25.3 37.9



#64
Tetrachloroethene
Concen: 48.611 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.006 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion:164 Resp: 53321
Ion Ratio Lower Upper
164 100
166 123.6 96.0 144.0
129 92.5 79.4 119.0
131 89.9 72.6 109.0









#65
Chlorobenzene
Concen: 48.105 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 163490
Ion Ratio Lower Upper
112 100
114 32.9 26.4 39.6



#66
1,1,1,2-Tetrachloroethane
Concen: 50.111 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.006 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion:131 Resp: 60221
Ion Ratio Lower Upper
131 100
133 95.2 46.9 140.8
119 63.5 31.7 95.1





#67

Ethyl Benzene

Concen: 49.007 ug/l

RT: 10.195 min Scan# 1493

Delta R.T. 0.006 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 287881

Ion Ratio Lower Upper

91 100

106 30.5 23.9 35.9



#68

m/p-Xylenes

Concen: 96.606 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX043270.D




Acq: 03 Oct 2024 18:17

Tgt Ion:106 Resp: 214462

Ion Ratio Lower Upper

106 100

91 211.5 167.0 250.6








#69
o-Xylene
Concen: 48.428 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 106831
Ion Ratio Lower Upper
106 100
91 224.5 111.6 334.8



#70
Styrene
Concen: 50.463 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion:104 Resp: 180408
Ion Ratio Lower Upper
104 100
78 55.3 43.5 65.3
103 54.9 43.9 65.9





#71

Bromoform

Concen: 50.991 ug/l

RT: 10.799 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX043270.D





Acq: 03 Oct 2024 18:17

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:173	Resp:	46942
Ion Ratio	Lower	Upper
173	100	
175	48.1	24.3 72.9
254	0.0	0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.018 min Scan# 1793


Delta R.T. -0.006 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17

Tgt Ion:152	Resp:	75329
Ion Ratio	Lower	Upper
152	100	
115	87.0	44.1 132.3
150	170.4	0.0 347.4









#73
Isopropylbenzene
Concen: 47.711 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:105 Resp: 279425
Ion Ratio Lower Upper
105 100
120 26.4 12.9 38.7



#74
N-ethyl acetate
Concen: 49.107 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 43 Resp: 136527
Ion Ratio Lower Upper
43 100
70 41.2 33.1 49.7
55 23.1 18.6 27.8
61 23.3 18.7 28.1









#75
1,1,2,2-Tetrachloroethane
Concen: 49.035 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 83 Resp: 93555
Ion Ratio Lower Upper
83 100
131 10.3 5.2 15.6
85 64.9 31.9 95.5



#76
1,2,3-Trichloropropane
Concen: 63.885 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 75 Resp: 106744
Ion Ratio Lower Upper
75 100
77 32.6 22.0 66.0





#77

Bromobenzene

Concen: 47.537 ug/l

RT: 11.195 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:156 Resp: 67752

Ion Ratio Lower Upper

156 100

77 163.0 82.0 245.8

158 96.4 48.1 144.3



#78

n-propylbenzene

Concen: 49.068 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. -0.000 min

Lab File: VX043270.D




Acq: 03 Oct 2024 18:17


Tgt Ion: 91 Resp: 315475

Ion Ratio Lower Upper

91 100

120 22.7 11.4 34.1









#79
2-Chlorotoluene
Concen: 47.065 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 194540
Ion Ratio Lower Upper
91 100
126 33.4 16.6 49.7



#80
1,3,5-Trimethylbenzene
Concen: 48.704 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 105 Resp: 235295
Ion Ratio Lower Upper
105 100
120 46.7 23.9 71.8









#81
trans-1,4-Dichloro-2-butene
Concen: 45.280 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 29193
Ion Ratio Lower Upper
75 100
53 112.1 83.9 125.9
89 49.4 38.6 58.0



#82
4-Chlorotoluene
Concen: 48.101 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 91 Resp: 226491
Ion Ratio Lower Upper
91 100
126 28.3 14.1 42.3





#83

tert-Butylbenzene

Concen: 47.868 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:119 Resp: 238251

Ion Ratio Lower Upper

119 100

91 63.1 31.0 93.0

134 23.2 11.5 34.4



#84

1,2,4-Trimethylbenzene

Concen: 48.122 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX043270.D




Acq: 03 Oct 2024 18:17


Tgt Ion:105 Resp: 233919

Ion Ratio Lower Upper

105 100


120 44.2 21.2 63.6








#85
sec-Butylbenzene
Concen: 48.800 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:105 Resp: 287933
Ion Ratio Lower Upper
105 100
134 19.8 10.1 30.1



#86
p-Isopropyltoluene
Concen: 48.657 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion:119 Resp: 239521
Ion Ratio Lower Upper
119 100
134 25.6 13.0 38.9
91 26.2 13.0 39.0





#87

1,3-Dichlorobenzene

Concen: 48.287 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min


Lab File: VX043270.D

Acq: 03 Oct 2024 18:17

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:146 Resp: 121233

Ion Ratio Lower Upper

146 100

111 42.2 21.8 65.3

148 64.0 32.4 97.2



#88

1,4-Dichlorobenzene

Concen: 46.577 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.006 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17




Tgt Ion:146 Resp: 120562


Ion Ratio Lower Upper

146 100

111 42.2 20.8 62.5


148 64.4 32.4 97.0








#89
n-Butylbenzene
Concen: 50.016 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 91 Resp: 206130
Ion Ratio Lower Upper
91 100
92 53.3 27.0 81.0
134 25.2 12.9 38.6



#90
Hexachloroethane
Concen: 48.507 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 117 Resp: 44782
Ion Ratio Lower Upper
117 100
201 65.5 34.0 101.8









#91
1,2-Dichlorobenzene
Concen: 47.691 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:146 Resp: 122452
Ion Ratio Lower Upper
146 100
111 44.9 21.9 65.5
148 64.4 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 47.920 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.006 min
Lab File: VX043270.D
Acq: 03 Oct 2024 18:17

Tgt Ion: 75 Resp: 22322
Ion Ratio Lower Upper
75 100
155 78.9 38.8 116.4
157 99.6 49.0 147.0





#93

1,2,4-Trichlorobenzene

Concen: 49.578 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX043270.D





Acq: 03 Oct 2024 18:17

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:	180	Resp:	73386
Ion Ratio	Lower	Upper	
180	100		
182	94.5	47.1	141.3
145	32.9	16.6	49.8



#94

Hexachlorobutadiene

Concen: 50.601 ug/l




RT: 13.725 min Scan# 2073


Delta R.T. -0.000 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17

Tgt Ion:	225	Resp:	29692
Ion Ratio	Lower	Upper	
225	100		
223	64.6	30.8	92.4
227	63.5	32.2	96.6





#95

Naphthalene

Concen: 48.276 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:128 Resp: 262695

Ion Ratio Lower Upper

128 100

127 12.7 10.2 15.2

129 10.9 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 48.189 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. -0.006 min

Lab File: VX043270.D

Acq: 03 Oct 2024 18:17

Tgt Ion:180 Resp: 73595

Ion Ratio Lower Upper

180 100

182 98.3 46.5 139.3

145 35.8 17.2 51.6

