

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
 Data File : VX043246.D
 Acq On : 03 Oct 2024 09:00
 Operator : JC/MD
 Sample : VX1003WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1003WBL01

Quant Time: Oct 04 01:24:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	87367	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	169225	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	150252	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	65184	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	74923	52.510	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.020%
35) Dibromofluoromethane	5.385	113	55843	47.850	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.700%
50) Toluene-d8	8.647	98	203106	50.220	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.440%
62) 4-Bromofluorobenzene	11.079	95	72998	49.683	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
Data File : VX043246.D
Acq On : 03 Oct 2024 09:00
Operator : JC/MD
Sample : VX1003WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1003WBL01

Quant Time: Oct 04 01:24:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

