

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
 Data File : VX043249.D
 Acq On : 03 Oct 2024 10:11
 Operator : JC/MD
 Sample : P4231-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D-20240924

Quant Time: Oct 04 01:27:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	80727	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	155132	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	135205	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57841	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72047	54.648	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.300%			
35) Dibromofluoromethane	5.385	113	52400	48.979	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.960%			
50) Toluene-d8	8.647	98	187641	50.611	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.220%			
62) 4-Bromofluorobenzene	11.079	95	65524	48.647	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 97.300%			
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.319	101	14673	16.245	ug/l	99
12) 1,1-Dichloroethene	2.313	96	2494	2.781	ug/l	97
27) cis-1,2-Dichloroethene	4.489	96	5211	4.681	ug/l	88
44) Trichloroethene	7.123	130	171455	157.749	ug/l	97
64) Tetrachloroethene	9.269	164	1235	1.316	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
Data File : VX043249.D
Acq On : 03 Oct 2024 10:11
Operator : JC/MD
Sample : P4231-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D-20240924

Quant Time: Oct 04 01:27:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

