

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
 Data File : VX043251.D
 Acq On : 03 Oct 2024 10:57
 Operator : JC/MD
 Sample : P4231-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW10-MW01D-20240926

Quant Time: Oct 04 01:28:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	84263	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	164836	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	141753	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59774	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	73528	53.430	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.860%
35) Dibromofluoromethane	5.385	113	54723	48.139	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.280%
50) Toluene-d8	8.647	98	195658	49.666	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.340%
62) 4-Bromofluorobenzene	11.079	95	68704	48.005	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.020%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	953	1.658	ug/l	97
44) Trichloroethene	7.135	130	1366	1.183	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
Data File : VX043251.D
Acq On : 03 Oct 2024 10:57
Operator : JC/MD
Sample : P4231-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW10-MW01D-20240926

Quant Time: Oct 04 01:28:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

