

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
 Data File : VX043252.D
 Acq On : 03 Oct 2024 11:21
 Operator : JC/MD
 Sample : P4231-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW10A-MW01I-20240926

Quant Time: Oct 04 01:28:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	86603	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	164220	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	145704	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58462	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	74326	52.551	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.100%
35) Dibromofluoromethane	5.385	113	55042	48.601	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.200%
50) Toluene-d8	8.647	98	198928	50.686	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.380%
62) 4-Bromofluorobenzene	11.079	95	68860	48.295	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.580%
Target Compounds						
16) Acetone	2.386	43	1069	1.810	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
Data File : VX043252.D
Acq On : 03 Oct 2024 11:21
Operator : JC/MD
Sample : P4231-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW10A-MW01I-20240926

Quant Time: Oct 04 01:28:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

