

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
 Data File : VX043253.D
 Acq On : 03 Oct 2024 11:44
 Operator : JC/MD
 Sample : P4231-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW10A-MW01S-20240926

Quant Time: Oct 04 01:29:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	84431	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	161707	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	138637	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54476	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	71211	51.644	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.280%
35) Dibromofluoromethane	5.385	113	53045	47.566	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.140%
50) Toluene-d8	8.647	98	193378	50.037	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.080%
62) 4-Bromofluorobenzene	11.079	95	65267	46.486	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.980%
Target Compounds						
16) Acetone	2.386	43	831	1.443	ug/l	90
22) Diisopropyl ether	3.770	45	7484	2.414	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
Data File : VX043253.D
Acq On : 03 Oct 2024 11:44
Operator : JC/MD
Sample : P4231-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW10A-MW01S-20240926

Quant Time: Oct 04 01:29:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

