

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
 Data File : VX043258.D
 Acq On : 03 Oct 2024 13:39
 Operator : JC/MD
 Sample : P4278-02 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BUR-1280

Quant Time: Oct 04 01:31:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	82675	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	161555	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	148645	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64992	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	73246	54.248	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.500%			
35) Dibromofluoromethane	5.385	113	53225	47.772	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.540%			
50) Toluene-d8	8.647	98	198036	51.291	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.580%			
62) 4-Bromofluorobenzene	11.079	95	74599	53.183	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 106.360%			

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
Data File : VX043258.D
Acq On : 03 Oct 2024 13:39
Operator : JC/MD
Sample : P4278-02 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BUR-1280

Quant Time: Oct 04 01:31:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

