

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
 Data File : VX043261.D
 Acq On : 03 Oct 2024 14:48
 Operator : JC/MD
 Sample : P4231-02DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D1-20240924DL

Quant Time: Oct 04 01:32:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	85008	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	162281	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	146400	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62906	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72284	52.066	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.140%			
35) Dibromofluoromethane	5.385	113	53781	48.055	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.120%			
50) Toluene-d8	8.647	98	197652	50.962	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.920%			
62) 4-Bromofluorobenzene	11.079	95	71678	50.872	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 101.740%			
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.325	101	2555	2.686	ug/l	95
12) 1,1-Dichloroethene	2.313	96	851	0.901	ug/l	96
44) Trichloroethene	7.123	130	42601	37.469	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
Data File : VX043261.D
Acq On : 03 Oct 2024 14:48
Operator : JC/MD
Sample : P4231-02DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D1-20240924DL

Quant Time: Oct 04 01:32:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

