

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
 Data File : VX043265.D
 Acq On : 03 Oct 2024 16:21
 Operator : JC/MD
 Sample : P4263-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Oct 04 01:34:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	85569	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	160528	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	141983	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63595	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72074	51.575	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.140%			
35) Dibromofluoromethane	5.385	113	53853	48.645	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.280%			
50) Toluene-d8	8.647	98	195065	50.845	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.680%			
62) 4-Bromofluorobenzene	11.079	95	70199	50.366	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 100.740%			
Target Compounds						
16) Acetone	2.392	43	973	1.667	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
Data File : VX043265.D
Acq On : 03 Oct 2024 16:21
Operator : JC/MD
Sample : P4263-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Oct 04 01:34:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

